



CERTIFICATE OF ANALYSIS

**TRIONE, Ninhydrin Reagent (2 Parts)
T200**

Product: TRIONE, Ninhydrin Reagent (2 Parts)

Part Number: T200

Lot Number: 108031

Production Date: 26 Aug 2011

Expiration Date: 26 Aug 2017

Primary Application: Sodium amino acid analysis

Acceptance Criteria:

1. TRIONE color without shades of brown or purple(Y/N)
2. L-aspartic acid to L-proline ratio ≥ 3.3 (Y/N)

4
4

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by:  **Date:** 20 Sep 2011

VP OPR
Title

ACCEPTED

PICKERING LABORATORIES, INC.
QUALITY ASSURANCE

Rev. 7, 3/25/2011