



CERTIFICATE OF ANALYSIS

TRIONE, Ninhydrin Reagent (2 Parts) T200

Product: TRIONE, Ninhydrin Reagent (2 Parts)

Part Number: T200

Lot Number: 110018

Production Date: 17 Oct 2011

Expiration Date: 13 Oct 2012

Primary Application: Amino acid analysis

Acceptance Criteria:

1. TRIONE color without shades of brown or purple(Y/N)
2. L-aspartic acid to L-proline ratio ≥ 3.0 (Y/N)

Y
Y

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by:

James Murphy
VP Ops
Title

Date: 31 Oct 2011

ACCEPTED

PICKERING LABORATORIES, INC.
QUALITY ASSURANCE