



CERTIFICATE OF ANALYSIS

TRIONE, Ninhydrin Reagent (2 Parts) T200

Product: TRIONE, Ninhydrin Reagent (2 Parts)

Part Number: T200

Lot Number:

112008

Production Date:

12 Dec 2011

Expiration Date:

08 Dec 2012

Primary Application: Amino acid analysis

Acceptance Criteria:

1. TRIONE color without shades of brown or purple (Y/N)
2. L-aspartic acid to L-proline ratio ≥ 3.0 (Y/N)

Y
Y

Storage Conditions: Dry at 8 to 25 °C

Sample: Aspartic acid and Proline mixture.

Analysis Certified by:

John Murphy
V.P. OPER.
Title

Date:

22 Dec 2011

ACCEPTED

PICKERING LABORATORIES, INC.
QUALITY ASSURANCE