



CERTIFICATE OF ANALYSIS

TRIONE Ninhydrin Reagent (2 Parts) T200

Product: TRIONE, Ninhydrin Reagent (2 Parts)

Part Number: T200

Lot Number: 2203016

Production Date: March 8, 2022

Expiration Date: March 8, 2023

Primary Application: Amino acid analysis

Acceptance Criteria:

1. TRIONE color without shades of brown or purple (Y/N)
2. L-aspartic acid to L-proline ratio ≥ 3.0 (Y/N)

Y

Y

Storage Conditions: Dry at 8 to 25 °C

Sample: Aspartic acid and Proline mixture.

Analysis Certified by: Jim Murphy

President

Title

Date: March 15, 2022