



CERTIFICATE OF ANALYSIS

URIPREP

UP100 (1800-0034)

Product: URIPREP
Part Number: UP100 (1800-0034)
Lot Number: 103015
Test Start Date: 08 MAR 2011

Primary Application: Prepare urine and other native samples for amino acid analysis.

Acceptance Criteria

Actual Values

1. pH is 1.29 ± 0.01 at 25°C
2. Retention times of Threonine and Serine are within ± 0.2 min of calibration standard

1.291
4

Calibration Standard:

Physiological Fluid Calibration Standard, PN 011006p, LN 909011

Analysis Certified by: *Jim Murphy* Date: 21 MAR 2011

VP OPR
Title

ACCEPTED
PICKERING LABORATORIES, INC.
QUALITY ASSURANCE