



CERTIFICATE OF ANALYSIS

URIPREP

UP100 (1800-0034)

Product: URIPREP
Part Number: UP100 (1800-0034)
Lot Number: 103015
Production Date: 08 MAR 2011
Expiration Date: 08 MAR 2012

Primary Application: Prepare urine and other native samples for amino acid analysis.

Acceptance Criteria

1. pH is 1.29 ± 0.01 at $25^{\circ} \text{C} \pm 0.4^{\circ} \text{C}$ (Y/N)
2. Retention times of Threonine and Serine are within ± 0.2 min of calibration standard (Y/N)
3. No peaks in the baseline (Y/N)

y

y

y

Calibration Standard:

Physiological Fluid Calibration Standard, PN 011006P

Analysis Certified by: _____ Date: 12 Apr 2011

John P. [Signature]
VIP Oper

Title

ACCEPTED
PICKERING LABORATORIES, INC.
QUALITY ASSURANCE

Rev. 7, 04/08/2011