



CERTIFICATE OF ANALYSIS

URIPREP

UP100 (1800-0034)

Product: URIPREP
Part Number: UP100 (1800-0034)
Lot Number: 107039
Production Date: 29 Jul 2011
Expiration Date: 28 Jul 2012

Primary Application: Prepare urine and other native samples for amino acid analysis.


Acceptance Criteria

1. pH is 1.29 ± 0.01 at $25^{\circ} \text{C} \pm 0.4^{\circ} \text{C}$ (Y/N)
2. Retention times of Threonine and Serine are within ± 0.2 min of calibration standard (Y/N)
3. No peaks in the baseline (Y/N)

Y
Y
Y

Calibration Standard:

Physiological Fluid Calibration Standard, PN 011006P

Analysis Certified by:  Date: 20 Sep 2011
VP DPR
Title

ACCEPTED

PICKERING LABORATORIES, INC.
QUALITY ASSURANCE

Rev. 7, 04/08/2011