



# CERTIFICATE OF ANALYSIS

URIPREP

UP100

Product: URIPREP  
Part Number: UP100  
Lot Number: 110021  
Production Date: 20 Oct 2011  
Expiration Date: 19 Oct 2012

**Primary Application:** Preparation of urine and other native samples for amino acids analysis.

### Acceptance Criteria

1. pH is  $1.29 \pm 0.01$  at  $25^\circ \text{C} \pm 0.4^\circ \text{C}$  (Y/N)
2. Retention times difference for Threonine and Serine meets specification (Y/N)
3. No peaks in the baseline (Y/N)

Y  
Y  
Y

**Sample:** Urine sample prepared with URIPREP

**Storage Conditions:** Dry at 8 to 25 °C

Analysis Certified by:

*[Signature]*  
VP OPR  
Title

Date:

31 Oct 2011

## ACCEPTED

PICKERING LABORATORIES, INC.  
QUALITY ASSURANCE