



CERTIFICATE OF ANALYSIS

URIPREP

UP100

Product: URIPREP

Part Number: UP100

Lot Number: 2303014

Production Date: 13 Mar 2023

Expiration Date: 13 Mar 2024

Primary Application: Preparation of urine and other native samples for amino acids analysis

Acceptance Criteria:

- | | |
|--|----------|
| 1. pH is 1.29 ± 0.01 at $25^{\circ} \text{C} \pm 0.4^{\circ}\text{C}$ (Y/N) | <u>Y</u> |
| 2. Retention times difference for Threonine and Serine meets specification (Y/N) | <u>Y</u> |
| 3. Baseline: no peaks (Y/N) | <u>Y</u> |

Sample: Urine sample prepared with URIPREP

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by: David Mazawa **Date:** 14 Mar 2023
Title: Technical Support Chemist