



CERTIFICATE OF ANALYSIS

URIPREP

UP100

Product: URIPREP

Part Number: UP100

Lot Number: 2501007

Production Date: 06 Jan 2025

Expiration Date: 06 Jan 2026

Primary Application: Preparation of urine and other native samples for amino acids analysis

Acceptance Criteria:

1. pH is 1.29 ± 0.01 at $25^{\circ} \text{C} \pm 0.4^{\circ}\text{C}$ (Y/N)	<u>Y</u>
2. Retention times difference for Threonine and Serine meets specification (Y/N)	<u>Y</u>
3. Baseline: no peaks (Y/N)	<u>Y</u>

Sample: Urine sample prepared with URIPREP

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by:	<u>David Mazawa</u>	Date:	<u>10 Jan 2025</u>
Title:	<u>QC Manager</u>		