



CERTIFICATE OF ANALYSIS

URIPREP

UP100

Product: URIPREP

Part Number: UP100

Lot Number: 2505046

Production Date: 22 May 2025

Expiration Date: 22 May 2026

Primary Application: Preparation of urine and other native samples for amino acids analysis

Acceptance Criteria:

1. pH is 1.29 ± 0.01 at $25^{\circ} \text{C} \pm 0.4^{\circ}\text{C}$ (Y/N)
2. Retention times difference for Threonine and Serine meets specification (Y/N)
3. Baseline: no peaks (Y/N)

Y

Y

Y

Sample: Urine sample prepared with URIPREP

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by: David Mazawa

Title: QC Manager

Date: 23 May 2025