



## CERTIFICATE OF ANALYSIS

**URIPREP**

**UP100**

**Product:** URIPREP

**Part Number:** UP100

**Lot Number:** 2510008

**Production Date:** 02 Oct 2025

**Expiration Date:** 02 Oct 2026

**Primary Application:** Preparation of urine and other native samples for amino acids analysis

### Acceptance Criteria:

- |                                                                                  |          |
|----------------------------------------------------------------------------------|----------|
| 1. pH is $1.29 \pm 0.01$ at $25^{\circ} \text{C} \pm 0.4^{\circ}\text{C}$ (Y/N)  | <u>Y</u> |
| 2. Retention times difference for Threonine and Serine meets specification (Y/N) | <u>Y</u> |
| 3. Baseline: no peaks (Y/N)                                                      | <u>Y</u> |

**Sample:** Urine sample prepared with URIPREP

**Storage Conditions:** Dry at 8 to 25 °C

|                               |                     |              |                    |
|-------------------------------|---------------------|--------------|--------------------|
| <b>Analysis Certified by:</b> | <u>David Mazawa</u> | <b>Date:</b> | <u>03 Oct 2025</u> |
| <b>Title:</b>                 | <u>QC Manager</u>   |              |                    |